

STEMprep™ Mouse Brain Dissociation Kit



For processing 50 brains using the STEMprep™ Tissue Dissociation System

Scientists Helping Scientists™ | WWW.STEMCELL.COM

Catalog #100-2130

50 Preparations

TOLL FREE PHONE 1 800 667 0322 • PHONE +1 604 877 0713

INFO@STEMCELL.COM • TECHSUPPORT@STEMCELL.COM

FOR GLOBAL CONTACT DETAILS VISIT OUR WEBSITE

Product Description

STEMprep™ Mouse Brain Dissociation Kit provides an optimized enzymatic dissociation cocktail for processing brain tissues into single-cell suspensions for studying microglia, oligodendrocytes, astrocytes, endothelial cells, as well as neural stem and progenitor cells. When used with the STEMprep™ Tissue Dissociator (Catalog #100-2112), mechanical dissociation is combined with enzymatic degradation of the extracellular matrix, efficiently breaking down tissue structure while preserving cellular integrity. Following dissociation, an optional density gradient centrifugation step can be performed to remove myelin prior to downstream use. The resulting single cells are immediately suitable for downstream applications, such as cell separation, culture, or various analyses.

For best results, use with STEMprep™ Tissue Dissociator and STEMprep™ Sample Tubes (Catalog #200-0800).

Product Information

The following components are sold as part of STEMprep™ Mouse Brain Dissociation Kit (Catalog #100-2130) and are not available for individual sale.

COMPONENT NAME	COMPONENT #	SIZE	STORAGE	SHELF LIFE
STEMprep™ Enzyme A	100-2131	1 x 2.5 mL	Store at -20°C.	Stable until expiry date (EXP) on label.
STEMprep™ Enzyme D	100-2133	1 x 1.25 mL	Store at 2 - 8°C. Do not freeze.	Stable until expiry date (EXP) on label.
STEMprep™ Enzyme Diluent Z	100-2134	1 x 125 mL	Store at 2 - 8°C. Do not freeze.	Stable until expiry date (EXP) on label.

Materials Required but Not Included

PRODUCT NAME	CATALOG #
50 mL Conical Tubes	e.g. 38010
70 µm Cell Strainer	e.g. 27260
Culture Dish, Non-Treated	e.g. 38045
D-PBS (Without Ca++ and Mg++)	37350
EasySep™ Buffer OR PBS containing 2% FBS and 1 mM EDTA	20144
STEMprep™ Sample Tubes	200-0800
STEMprep™ Tissue Dissociator	100-2112

Preparation of Reagents and Materials

Thaw STEMprep™ Enzyme A at room temperature (15 - 25°C) for immediate use or overnight at 2 - 8°C. Do not thaw at 37°C.

NOTE: Once thawed, use immediately or aliquot and store at -20°C until the expiry date as indicated on the label. After thawing the aliquots, use immediately. Do not re-freeze.

BRAIN DISSOCIATION ENZYME COCKTAIL

Prepare brain dissociation enzyme cocktail fresh before use.

For one brain, prepare 2.5 mL of brain dissociation enzyme cocktail by combining the following in a STEMprep™ Sample Tube:

- 2.425 mL of Enzyme Diluent Z
- 50 µL of Enzyme A
- 25 µL of Enzyme D

Sample Preparation

Dissect mouse brain into a non-treated culture dish containing cold PBS.

Directions for Use

1. Transfer the brain tissue (325 - 395 mg) into the STEMprep™ Sample Tube containing 2.5 mL of brain dissociation enzyme cocktail.
NOTE: Plastic Rack for Centrifuge Tubes, 50 mL (Catalog #200-0651) is provided with each STEMprep™ Tissue Dissociator System to manage STEMprep™ Sample Tubes while loading or removing tissues.
2. Tightly close the STEMprep™ Sample Tube lid (ensuring proper closure) and insert it into a slot on the STEMprep™ Tissue Dissociator. For complete instructions on using the instrument, refer to the Technical Manual: STEMprep™ Tissue Dissociator (Document #10000030598), available at www.stemcell.com, or contact us to request a copy.
NOTE: For best results, avoid pinning the tissue with the rotor in the center of the STEMprep™ Sample Tube. Ensure total sample volume containing tissue does not exceed the "MAX" fill line on the tube.
3. Select and run the "Mouse Brain" dissociation protocol on the STEMprep™ Tissue Dissociator.
NOTE: If the tube will not be removed immediately after protocol completion, add a 4°C hold to preserve sample integrity.
4. After the protocol is complete, remove the STEMprep™ Sample Tube from the slot. Place a 70 µm cell strainer in a 50 mL conical tube. Pre-wet the strainer with 5 mL of EasySep™ Buffer. Carefully pour the sample over the strainer.
5. Rinse the STEMprep™ Sample Tube with 10 mL of cold EasySep™ Buffer and pour the wash over the strainer. Repeat this wash step once more. Top up the conical tube to 50 mL with EasySep™ Buffer and screw on the cap. Discard the strainer and STEMprep™ Sample Tube.
NOTE: Small tissue fragments may remain in the filter after dissociation; these typically do not affect yield. Larger tissue pieces can be gently pushed through the strainer using the rubber end of a syringe plunger.
6. Centrifuge the conical tube at 300 x g for 10 minutes at room temperature with the brake on low. After centrifugation, carefully aspirate the supernatant.
7. OPTIONAL: If desired, perform red blood cell lysis by adding 5 mL of Ammonium Chloride Solution (Catalog #07850) to the cell pellet. Thoroughly mix the cell suspension by pipetting up and down. Incubate on ice for 10 minutes. Top up sample to 50 mL with EasySep™ Buffer. Centrifuge at 300 x g for 10 minutes at room temperature with the brake on low. After centrifugation, carefully aspirate the supernatant.
NOTE: If proceeding to EasySep™, do not perform red blood cell lysis.
8. OPTIONAL: If desired, perform myelin removal with OptiPrep™ (Catalog #07820). See below for instructions.
 - a. Prepare 17 mL of 18% v/v OptiPrep™ solution per sample by adding 3.1 mL OptiPrep™ to 13.9 mL EasySep™ Buffer. Gently invert the tube to mix, do not shake or vortex. Store at room temperature until use.
 - b. Add 17 mL of 18% OptiPrep™ solution to the cell pellet and gently pipette up and down. Avoid bubbles and foam.
 - c. Gently overlay 3 mL of cold EasySep™ Buffer on top of the cell suspension.
NOTE: Slowly add EasySep™ Buffer down the wall of the tube without disturbing the interface.
 - d. Centrifuge at 1300 x g for 15 minutes at room temperature with no brake.
 - e. Slowly aspirate and discard the top layer of EasySep™ Buffer, the myelin layer, and OptiPrep™ (Figure 1).
 - f. Top up the tube to 50 mL with EasySep™ Buffer and resuspend the cell pellet. Centrifuge at 300 x g for 10 minutes. Carefully aspirate the supernatant.
9. Resuspend cells in the desired volume of EasySep™ Buffer or medium of your choice. Cells are ready for downstream use.




Figure 1. Myelin Removal with OptiPrep™

Data




Figure 2. Cell Viability and Yield Obtained Using the STEMprep™ Mouse Brain Dissociation Kit

Brains from healthy C57BL/6 mice were dissociated into single-cell suspensions using the STEMprep™ Mouse Brain Dissociation Kit on the STEMprep™ Tissue Dissociator. Following STEMprep™ processing and myelin removal, viability and yield of single-cell suspensions were assessed by flow cytometry. Data are presented as mean ± SD.

Related Products

For more information about STEMprep™ kits and protocols, visit www.stemcell.com/stemprep, contact us at techsupport@stemcell.com.

PRODUCTS ARE FOR RESEARCH USE ONLY AND NOT INTENDED FOR HUMAN OR ANIMAL DIAGNOSTIC OR THERAPEUTIC USES UNLESS OTHERWISE STATED. FOR ADDITIONAL INFORMATION ON QUALITY AT STEMCELL, REFER TO WWW.STEMCELL.COM/COMPLIANCE.

Copyright © 2025 by STEMCELL Technologies Inc. All rights reserved including graphics and images. STEMCELL Technologies & Design, STEMCELL Shield Design, Scientists Helping Scientists, and STEMprep are trademarks of STEMCELL Technologies Canada Inc. OptiPrep is a trademark of Serumwerk Bernburg AG. The products sold under the OptiPrep brand name are also manufactured by Serumwerk Bernburg AG. All other trademarks are the property of their respective holders. While STEMCELL has made all reasonable efforts to ensure that the information provided by STEMCELL and its suppliers is correct, it makes no warranties or representations as to the accuracy or completeness of such information.